

Video-based and interference-free force analysis

Sebastian Knust^{1*}, Andy Sischka¹, Andre Spiering¹, Dario Anselmetti¹

¹Experimental Biophysics and Applied Nanoscience, Bielefeld University, 33615 Bielefeld, Germany

Experimental Biophysics Group

ELEFELD UNIVERSITY INTER FOR BIOTECHNOLOGY (CEBITEC) ELEFELD INSTITUTE FOR BIOPHYSICS & NANOSCIENCE (BINAS)

* sknust@physik.uni-bielefeld.de

BP 8.17

Abstract

We have developed a video-based axial force detection and analysis system for optical tweezers for measuring the minute forces exerted on single molecules during controlled translocation through nanopores with sub-piconewton precision [1].

Because of the limited sample rate of the video signal of 123 Hz, we integrated Allan variance analysis [2] for trap stiffness calibration.

When manipulating a microbead in the vicinity of a weakly reflecting surface with simultaneous axial force detection, interference effects have to be considered [3]. Therefore, we measured and analyzed the backscattering light properties of polystyrene and silica microbeads with different diameters and propose distinct and optimized experimental configurations (microbead material and diameter) for minimal light backscattering and virtually interference-free microbead position detection. As a proof of principle, we investigated the nanopore threading forces of a single dsDNA strand attached to a trapped microbead with an overall force resolution of 0.5 pN.

Experimental Setup



Bead size and interference

When working in the vicinity of surfaces, even weak reflections lead to unwanted interference effects. Some beads however reflect only a small fraction of light. They therefore exhibit only negligible interference, however backscattering is limited as well. Only video-based force analysis is capable of measuring low-noise interference-free signals.



Comparison of interference effects

Sciences

on Optical Tw

5th International Symposium

For illumination, we use a Schott KL2500 LED cold light source with a maximum output of 1000 lm.

A custom built post-magnification with a magnifying factor of 10 and an AVT Guppy Pro F-031 monochrome firewire CCD camera with 14-bit image acquisition at 123 fps are utilized for image acquisition.



With the pixel size of 5.6 μ m, we get a conversion factor of approximately 9 nm/px. Thus, a 3.05 μ m bead has an apparent radius in the range of 170 px.

A 1pN force corresponds to a bead radius change of approximately 0.8 px at 250 mW laser power.



Size Detection

We manually select circular regions of interest for rising (green) and falling (red) edges. The strongest rising and falling edges (green and red points) are detected along radial spokes (blue). The orange dots are the resulting mid-points between the two edges, through which the yellow circle is fitted.

For clarity, only every tenth spoke with corresponding edges is shown.

Controlled dsDNA threading through 55nm NP

 $3.04 \ \mu m$ bead in vicinity of Si₃N₄ nanopore with applied voltage of 50 mV, laser power 250 mW



Force calibration

Allan variance (AV) allows for easy calibration of low frequency signals. It is defined as half the variance of the averaged difference between two consecutive local averaged position samples: $\sigma^{2}(\tau) = \frac{1}{2} \left\langle (\bar{x}_{\tau,j+1} - \bar{x}_{\tau,j})^{2} \right\rangle \qquad \bar{x}_{\tau,j} = \frac{1}{\tau} \int_{\tau(j-0.5)}^{\tau(j+0.5)} dt \, x(t)$



τ [S]

The AV of a trapped bead, which is a damped harmonic oscillator, can be analytically expressed [2] as



Conclusion

Our high precision setup delivers *interference-free* axial force detection with a resolution of ±0.5 pN at a sample rate of 123 Hz. Lateral force detection can be easily implemented as well.

 $0.001\tau_{c}$

References

[1] S. Knust et. al., Rev. Sci. Instrum. 83, 103704 (2012)
[2] B. Lansdorp, O. Saleh, Rev. Sci. Instrum. 83, 025115 (2012)
[3] A. Sischka et. al., Rev. Sci. Instrum. 79, 63702 (2008)

Supported by:





